

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PURINGTON,PATRICK,A
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: PURINGTON,PATRICK
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 05/22/2019

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24818080A2
DRY GAS EXPIRY DATE: 10/05/2020
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: WAASER

DATE/TIME OF TEST: 05/25/2019 04:18:02

Diagnostics Check	OK	04:11
Air Blank	0.000	04:11
Control Test	0.080	04:12
Air Blank	0.000	04:12
Subject Sample # 1	0.125	04:13
Breath Volume	1.457	04:13
Air Blank	0.000	04:13
Air Blank	0.000	04:15
Subject Sample # 2	0.120	04:16
Breath Volume	1.871	04:16
Air Blank	0.000	04:17
Control Test	0.080	04:17
Air Blank	0.000	04:17
Diagnostics Check	OK	04:18

Data Download Date/Time: 06/28/2019 14:41